

LIBERO W All Data Report

Additional Information

Download the LIBERO software from www.elpro.com/downloads

- Use liberoCONFIG to configure LIBERO with your own settings
- Use elproVIEWER to access all recorded data and create own reports

Device Information

Base Type:	LIBERO WB V1.11	Sensor Type:	WSI V1.10	Status:	Logging
Base ID:	81000001	Sensor ID:	82000002	Log Interval:	1 m
Base Battery State:	Low	Sensor Battery State:	Good	Report Time Base:	UTC +01:00
Profile ID / Checksum:	A001 / 2.498.997.501	Start at:	03.May.2016 09:39:44	Calibration Expiry Date:	Not available
Configured by:	EC 117 / tvogel, 26.Sep.2016 0037:07				

Report Information

From:	03.May.2016 09:39:44	First Warning at:	none	Highest Temperature:	7.6 °C; 24.Jun.2016 04:29:44
To:	23.Aug.2016 11:40:55	First Alarm at:	none	Lowest Temperature:	2.2 °C; 05.Jul.2016 44:49:44
Days in Report:	113	MKT:	5.5 °C	Average Temperature:	5.4 °C

